

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120924\
 Data File : VU062158.D
 Acq On : 09 Dec 2024 14:45
 Operator : MD/SY
 Sample : P5116-06
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E1PK0

Quant Time: Dec 09 15:57:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Dec 09 13:19:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	105817	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	99333	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	47749	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	30133	4.222	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	84.400%
7) Chloroethane-d5	1.907	69	22901	4.511	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.200%
11) 1,1-Dichloroethene-d2	2.560	65	15004	4.419	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.400%
20) 2-Butanone-d5	4.621	46	71856	52.531	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.060%
24) Chloroform-d	5.052	84	70748	4.580	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
26) 1,2-Dichloroethane-d4	5.695	65	38167	5.025	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.400%
32) Benzene-d6	5.717	84	134370	4.697	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
36) 1,2-Dichloropropane-d6	6.682	67	40552	4.954	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.000%
41) Toluene-d8	7.891	98	123395	4.566	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.400%
43) trans-1,3-Dichloroprop...	8.177	79	14298	3.968	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.400%
46) 2-Hexanone-d5	8.627	63	54262	49.918	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.840%
56) 1,1,2,2-Tetrachloroeth...	10.749	84	32686	5.131	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	44875	5.100	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	102.000%
Target Compounds						
					Qvalue	
30) Cyclohexane	5.373	56	1121	0.099	ug/L	98
35) Methylcyclohexane	6.756	83	1639	0.127	ug/L	89
38) Bromodichloromethane	7.103	83	1540	0.146	ug/L #	67
42) Toluene	7.965	91	8130	0.245	ug/L	89
49) Dibromochloromethane	8.801	129	2192	0.321	ug/L	94
53) m,p-Xylene	9.698	106	1435	0.104	ug/L	96
59) Bromoform	10.286	173	760	0.203	ug/L #	93
63) 1,2,4-Trimethylbenzene	11.466	105	1930	0.067	ug/L	93
72) 1,2,3-Trichlorobenzene	14.328	180	1608	0.196	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120924\
Data File : VU062158.D
Acq On : 09 Dec 2024 14:45
Operator : MD/SY
Sample : P5116-06
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E1PK0

Quant Time: Dec 09 15:57:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Dec 09 13:19:03 2024
Response via : Initial Calibration

