

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120924\
 Data File : VU062185.D
 Acq On : 10 Dec 2024 03:12
 Operator : MD/SY
 Sample : P5109-17
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHJA6

Quant Time: Dec 10 03:50:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 10 00:32:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	99912	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.409	117	92192	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.804	152	44282	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	32061	4.757	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	95.200%
7) Chloroethane-d5	1.908	69	24161	5.040	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.800%
11) 1,1-Dichloroethene-d2	2.560	65	15853	4.945	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.800%
20) 2-Butanone-d5	4.624	46	64922	50.267	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.540%
24) Chloroform-d	5.052	84	73734	5.055	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.200%
26) 1,2-Dichloroethane-d4	5.692	65	38368	5.350	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.000%
32) Benzene-d6	5.718	84	139668	5.260	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.200%
36) 1,2-Dichloropropane-d6	6.682	67	40280	5.302	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.000%
41) Toluene-d8	7.891	98	126936	5.061	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
43) trans-1,3-Dichloroprop...	8.174	79	15895	4.753	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.000%
46) 2-Hexanone-d5	8.627	63	49947	49.507	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.020%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	31064	5.255	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	44278	5.426	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120924\
Data File : VU062185.D
Acq On : 10 Dec 2024 03:12
Operator : MD/SY
Sample : P5109-17
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHJA6

Quant Time: Dec 10 03:50:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 10 00:32:19 2024
Response via : Initial Calibration

