

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120924\
 Data File : VU062170.D
 Acq On : 09 Dec 2024 20:13
 Operator : MD/SY
 Sample : P5107-13
 Misc : 25mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BHJ90

Quant Time: Dec 10 00:35:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 10 00:32:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.242	114	102716	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	95265	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	43427	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.592	65	18066	2.608	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	52.200%
7) Chloroethane-d5	1.907	69	16801	3.409	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.200%
11) 1,1-Dichloroethene-d2	2.560	65	9204	2.792	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.800%#
20) 2-Butanone-d5	4.624	46	63005	47.451	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.900%
24) Chloroform-d	5.052	84	61251	4.085	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.600%
26) 1,2-Dichloroethane-d4	5.695	65	32168	4.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
32) Benzene-d6	5.717	84	97123	3.540	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	70.800%
36) 1,2-Dichloropropane-d6	6.682	67	33207	4.230	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.600%
41) Toluene-d8	7.891	98	82415	3.180	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	63.600%#
43) trans-1,3-Dichloroprop...	8.177	79	10217	2.957	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	59.200%
46) 2-Hexanone-d5	8.627	63	42408	40.679	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.360%
56) 1,1,2,2-Tetrachloroeth...	10.746	84	28911	4.733	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	35583	4.446	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.000%
Target Compounds						Qvalue
34) Trichloroethene	6.537	95	8949	1.070	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU120924\
Data File : VU062170.D
Acq On : 09 Dec 2024 20:13
Operator : MD/SY
Sample : P5107-13
Misc : 25mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BHJ90

Quant Time: Dec 10 00:35:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR120924WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 10 00:32:19 2024
Response via : Initial Calibration





#34
 Trichloroethene
 Concen: 1.070 ug/L
 RT: 6.537 min Scan# 1
 Delta R.T. 0.003 min
 Lab File: VU062170.D
 Acq: 09 Dec 2024 20:13

Instrument :
 MSVOA_U
 ClientSampleId :
 BHJ90

Tgt Ion: 95 Resp: 8949

Ion	Ratio	Lower	Upper
95	100		
97	64.1	45.9	85.3
132	96.1	63.8	118.4
130	100.9	65.5	121.7

