

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U121019\
 Data File : VU036107.D
 Acq On : 11 Dec 2019 01:43
 Operator : JC/MD
 Sample : K6240-07
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFQY7

Quant Time: Dec 11 08:43:52 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Dec 11 07:30:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	260381	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	296054	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.84	152	99106	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	120416	5.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.80%
7) Chloroethane-d5	1.93	69	115818	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.00%
11) 1,1-Dichloroethene-d2	2.59	63	129694	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	4.70	46	244156	61.74	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	123.48%
24) Chloroform-d	5.11	84	186035	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
26) 1,2-Dichloroethane-d4	5.75	65	99691	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
32) Benzene-d6	5.77	84	352500	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.00%
36) 1,2-Dichloropropane-d6	6.73	67	114085	5.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.00%
41) Toluene-d8	7.93	98	282821	4.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.80%
43) trans-1,3-Dichloropropene-	8.21	79	41509	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
46) 2-Hexanone-d5	8.68	63	189562	55.62	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.24%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	107926	5.19	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.80%
64) 1,2-Dichlorobenzene-d4	12.22	152	98370	5.16	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.20%

Target Compounds					Ovalue
13) Acetone	2.68	43	9200	1.833 ug/L	91
16) Methylene chloride	3.08	84	2614	0.099 ug/L #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121019\
Data File : VU036107.D
Acq On : 11 Dec 2019 01:43
Operator : JC/MD
Sample : K6240-07
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFQY7

Quant Time: Dec 11 08:43:52 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Dec 11 07:30:58 2019
Response via : Initial Calibration

