

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121020\
 Data File : VU041537.D
 Acq On : 10 Dec 2020 15:31
 Operator : SY/MD
 Sample : L4995-14
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AS3

Quant Time: Dec 11 06:48:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 11 06:06:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	43758	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	37052	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	19717	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	7334	3.26	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.20%
7) Chloroethane-d5	1.92	69	8181	4.15	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	83.00%
11) 1,1-Dichloroethene-d2	2.58	63	15778	2.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.00%#
20) 2-Butanone-d5	4.66	46	35812	42.49	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.98%
24) Chloroform-d	5.08	84	31097	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
26) 1,2-Dichloroethane-d4	5.72	65	20939	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
32) Benzene-d6	5.74	84	44577	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.00%
36) 1,2-Dichloropropane-d6	6.70	67	12399	5.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.80%
41) Toluene-d8	7.90	98	39285	4.86	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.20%
43) trans-1,3-Dichloropropene-	8.19	79	7049	4.93	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.60%
46) 2-Hexanone-d5	8.64	63	24358	40.93	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.86%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	11490	4.32	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.40%
64) 1,2-Dichlorobenzene-d4	12.19	152	17788	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121020\
Data File : VU041537.D
Acq On : 10 Dec 2020 15:31
Operator : SY/MD
Sample : L4995-14
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AS3

Quant Time: Dec 11 06:48:40 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 11 06:06:19 2020
Response via : Initial Calibration

