

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121020\  
 Data File : VU041540.D  
 Acq On : 10 Dec 2020 17:03  
 Operator : SY/MD  
 Sample : L4995-17  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 VHBLK01

Quant Time: Dec 11 06:56:53 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR120420WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Fri Dec 11 06:06:19 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 6.26  | 114  | 37758    | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.42  | 117  | 36537    | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.81 | 152  | 19322    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.60   | 65    | 10336    | 5.33     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 106.60% |
| 7) Chloroethane-d5             | 1.92   | 69    | 9831     | 5.78     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 115.60% |
| 11) 1,1-Dichloroethene-d2      | 2.58   | 63    | 18954    | 4.04     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 80.80%  |
| 20) 2-Butanone-d5              | 4.66   | 46    | 27996    | 38.50    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 77.00%  |
| 24) Chloroform-d               | 5.08   | 84    | 26487    | 4.94     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 98.80%  |
| 26) 1,2-Dichloroethane-d4      | 5.72   | 65    | 18933    | 5.26     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 105.20% |
| 32) Benzene-d6                 | 5.74   | 84    | 40562    | 4.98     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 99.60%  |
| 36) 1,2-Dichloropropane-d6     | 6.70   | 67    | 12107    | 5.09     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 101.80% |
| 41) Toluene-d8                 | 7.90   | 98    | 39629    | 4.97     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 99.40%  |
| 43) trans-1,3-Dichloropropene- | 8.19   | 79    | 6704     | 4.75     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 95.00%  |
| 46) 2-Hexanone-d5              | 8.64   | 63    | 21086    | 35.93    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 71.86%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.75  | 84    | 10459    | 3.99     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 79.80%  |
| 64) 1,2-Dichlorobenzene-d4     | 12.19  | 152   | 17968    | 5.44     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 108.80% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121020\  
Data File : VU041540.D  
Acq On : 10 Dec 2020 17:03  
Operator : SY/MD  
Sample : L4995-17  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
VHBLK01

Quant Time: Dec 11 06:56:53 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR120420WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Fri Dec 11 06:06:19 2020  
Response via : Initial Calibration

