

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121020\
 Data File : VU041546.D
 Acq On : 10 Dec 2020 19:22
 Operator : SY/MD
 Sample : L5023-06
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0AW2

Quant Time: Dec 11 07:15:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 11 06:06:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	41731	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	35971	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	18554	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	6914	3.23	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	64.60%
7) Chloroethane-d5	1.92	69	8114	4.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.20%
11) 1,1-Dichloroethene-d2	2.58	63	15091	2.91	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.20%#
20) 2-Butanone-d5	4.66	46	35935	44.71	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.42%
24) Chloroform-d	5.08	84	29651	5.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.20%
26) 1,2-Dichloroethane-d4	5.72	65	19835	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.80%
32) Benzene-d6	5.74	84	41483	5.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.60%
36) 1,2-Dichloropropane-d6	6.70	67	11450	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	7.91	98	38818	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
43) trans-1,3-Dichloropropene-	8.19	79	7351	5.29	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	105.80%
46) 2-Hexanone-d5	8.64	63	27427	47.47	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.94%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	10804	4.18	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	17833	5.62	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121020\
Data File : VU041546.D
Acq On : 10 Dec 2020 19:22
Operator : SY/MD
Sample : L5023-06
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0AW2

Quant Time: Dec 11 07:15:36 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR120420WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 11 06:06:19 2020
Response via : Initial Calibration

