

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121021\
 Data File : VU046270.D
 Acq On : 10 Dec 2021 17:08
 Operator : SY/MD
 Sample : M4943-08
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGKW8

Quant Time: Dec 10 22:25:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 03 05:08:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	225510	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	232249	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	119808	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	69300	37.305	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.620%
7) Chloroethane-d5	1.919	69	54197	37.999	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	76.000%
11) 1,1-Dichloroethene-d2	2.575	63	99885	29.952	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.900%#
21) 2-Butanone-d5	4.636	46	135492	91.751	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.750%
24) Chloroform-d	5.067	84	135251	43.599	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.200%
26) 1,2-Dichloroethane-d4	5.707	65	93867	45.087	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.180%
32) Benzene-d6	5.732	84	285469	42.885	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.760%
36) 1,2-Dichloropropane-d6	6.694	67	90635	43.933	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.860%
41) Toluene-d8	7.899	98	251471	41.491	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.980%
43) trans-1,3-Dichloroprop...	8.179	79	45240	45.427	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.860%
47) 2-Hexanone-d5	8.636	63	93911	96.009	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	96.010%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	142069	45.409	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.820%
66) 1,2-Dichlorobenzene-d4	12.195	152	105641	45.786	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.580%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121021\
Data File : VU046270.D
Acq On : 10 Dec 2021 17:08
Operator : SY/MD
Sample : M4943-08
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGKW8

Quant Time: Dec 10 22:25:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 03 05:08:36 2021
Response via : Initial Calibration

