

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121118\
 Data File : VU028573.D
 Acq On : 11 Dec 2018 00:02
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 11 04:34:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.670	0.652	2.7	77	0.00
3 P	Chloromethane	1.232	1.208	1.9	81	0.00
4 C	Vinyl Chloride	0.889	0.893	-0.4#	80	0.00
5 T	Bromomethane	0.358	0.470	-31.3#	113	-0.02
6 T	Chloroethane	0.503	0.503	0.0	78	-0.01
7 T	Trichlorofluoromethane	0.936	0.918	1.9	80	0.00
8 T	Diethyl Ether	0.438	0.417	4.8	79	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.560	0.543	3.0	79	0.00
10 T	Methyl Iodide	0.439	0.482	-9.8	86	0.00
11 T	Tert butyl alcohol	0.170	0.203	-19.4	95	0.00
12 CM	1,1-Dichloroethene	0.557	0.536	3.8#	78	0.00
13 T	Acrolein	0.136	0.139	-2.2	81	0.00
14 T	Allyl chloride	1.317	1.450	-10.1	92	0.00
15 T	Acrylonitrile	0.472	0.517	-9.5	86	0.00
16 T	Acetone	0.498	0.421	15.5	72	0.00
17 T	Carbon Disulfide	2.005	1.913	4.6	79	0.00
18 T	Methyl Acetate	1.291	1.462	-13.2	89	0.00
19 T	Methyl tert-butyl Ether	2.110	2.144	-1.6	81	0.00
20 T	Methylene Chloride	0.734	0.690	6.0	81	0.00
21 T	trans-1,2-Dichloroethene	0.627	0.606	3.3	78	0.00
22 T	Diisopropyl ether	2.561	2.712	-5.9	84	0.00
23 T	Vinyl Acetate	2.112	2.246	-6.3	83	0.00
24 P	1,1-Dichloroethane	1.341	1.350	-0.7	82	0.00
25 T	2-Butanone	0.680	0.712	-4.7	82	0.00
26 T	2,2-Dichloropropane	1.062	0.948	10.7	73	0.00
27 T	cis-1,2-Dichloroethene	0.709	0.712	-0.4	81	0.00
28 T	Bromochloromethane	0.609	0.700	-14.9	91	0.00
29 T	Tetrahydrofuran	0.437	0.488	-11.7	87	0.00
30 C	Chloroform	1.168	1.202	-2.9#	83	0.00
31 T	Cyclohexane	1.573	1.314	16.5	81	0.00
32 T	1,1,1-Trichloroethane	0.970	0.988	-1.9	82	0.00
33 S	1,2-Dichloroethane-d4	0.763	0.822	-7.7	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.315	0.343	-8.9	89	0.00
36 T	1,1-Dichloropropene	0.547	0.542	0.9	80	0.00
37 T	Ethyl Acetate	0.707	0.770	-8.9	85	0.00
38 T	Carbon Tetrachloride	0.441	0.445	-0.9	83	0.00
39 T	Methylcyclohexane	0.668	0.623	6.7	77	0.00
40 TM	Benzene	1.611	1.634	-1.4	82	0.00
41 T	Methacrylonitrile	0.387	0.429	-10.9	87	0.00
42 TM	1,2-Dichloroethane	0.564	0.569	-0.9	82	0.00
43 T	Isopropyl Acetate	1.119	1.211	-8.2	85	0.00
44 TM	Trichloroethene	0.358	0.363	-1.4	83	0.00
45 C	1,2-Dichloropropane	0.453	0.470	-3.8#	81	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121118\
 Data File : VU028573.D
 Acq On : 11 Dec 2018 00:02
 Operator : MD/SY
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_U
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 11 04:34:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.269	0.271	-0.7	81	0.00
47 T	Bromodichloromethane	0.523	0.535	-2.3	83	0.00
48 T	Methyl methacrylate	0.556	0.616	-10.8	86	0.00
49 T	1,4-Dioxane	0.009	0.011	-22.2#	104	0.00
50 S	Toluene-d8	1.245	1.351	-8.5	89	0.00
51 T	4-Methyl-2-Pentanone	0.706	0.781	-10.6	88	0.00
52 CM	Toluene	0.948	0.955	-0.7#	80	0.00
53 T	t-1,3-Dichloropropene	0.623	0.631	-1.3	81	0.00
54 T	cis-1,3-Dichloropropene	0.676	0.667	1.3	79	0.00
55 T	1,1,2-Trichloroethane	0.375	0.383	-2.1	80	0.00
56 T	Ethyl methacrylate	0.642	0.667	-3.9	81	0.00
57 T	1,3-Dichloropropane	0.696	0.712	-2.3	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.329	0.330	-0.3	79	0.00
59 T	2-Hexanone	0.559	0.603	-7.9	86	0.00
60 T	Dibromochloromethane	0.367	0.369	-0.5	82	0.00
61 T	1,2-Dibromoethane	0.393	0.411	-4.6	82	0.00
62 S	4-Bromofluorobenzene	0.465	0.496	-6.7	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	80	0.00
64 T	Tetrachloroethene	0.334	0.334	0.0	82	0.00
65 PM	Chlorobenzene	1.065	1.047	1.7	80	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.356	-4.7	82	0.00
67 C	Ethyl Benzene	1.959	1.961	-0.1#	80	0.00
68 T	m/p-Xylenes	0.713	0.719	-0.8	79	0.00
69 T	o-Xylene	0.693	0.706	-1.9	80	0.00
70 T	Styrene	1.146	1.173	-2.4	81	0.00
71 P	Bromoform	0.289	0.312	-8.0	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.837	3.830	0.2	82	0.00
74 T	N-amyl acetate	2.213	2.355	-6.4	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.474	1.576	-6.9	86	0.00
76 T	1,2,3-Trichloropropane	1.277	1.480	-15.9	91	0.00
77 T	Bromobenzene	0.897	0.890	0.8	80	0.00
78 T	n-propylbenzene	4.767	4.688	1.7	80	0.00
79 T	2-Chlorotoluene	2.746	2.775	-1.1	82	0.00
80 T	1,3,5-Trimethylbenzene	3.176	3.207	-1.0	81	0.00
81 T	trans-1,4-Dichloro-2-butene	0.550	0.580	-5.5	98	0.00
82 T	4-Chlorotoluene	3.254	3.213	1.3	81	0.00
83 T	tert-Butylbenzene	3.061	3.044	0.6	81	0.00
84 T	1,2,4-Trimethylbenzene	3.273	3.280	-0.2	81	0.00
85 T	sec-Butylbenzene	3.933	3.901	0.8	80	0.00
86 T	p-Isopropyltoluene	3.341	3.253	2.6	79	0.00
87 T	1,3-Dichlorobenzene	1.694	1.648	2.7	81	0.00
88 T	1,4-Dichlorobenzene	1.753	1.659	5.4	80	0.00
89 T	n-Butylbenzene	3.381	3.210	5.1	76	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121118\
Data File : VU028573.D
Acq On : 11 Dec 2018 00:02
Operator : MD/SY
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC050

Quant Time: Dec 11 04:34:45 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.505	0.500	1.0	80	0.00
91 T	1,2-Dichlorobenzene	1.695	1.656	2.3	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.319	0.344	-7.8	84	0.00
93 T	1,2,4-Trichlorobenzene	1.215	1.096	9.8	78	0.00
94 T	Hexachlorobutadiene	0.555	0.493	11.2	75	0.00
95 T	Naphthalene	4.067	3.896	4.2	79	0.00
96 T	1,2,3-Trichlorobenzene	1.222	1.117	8.6	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6