

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121123\
 Data File : VU056737.D
 Acq On : 11 Dec 2023 17:17
 Operator : MD/SY
 Sample : 05728-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 12 01:11:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 12 01:07:35 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	253758	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	259349	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	108824	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	105956	4.871	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	97.400%	
7) Chloroethane-d5	1.907	69	81868	5.030	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	100.600%	
11) 1,1-Dichloroethene-d2	2.563	65	46488	4.995	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	100.000%	
20) 2-Butanone-d5	4.618	46	218378	50.838	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	101.680%	
24) Chloroform-d	5.055	84	202517	4.917	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.400%	
26) 1,2-Dichloroethane-d4	5.695	65	99392	5.099	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.000%	
32) Benzene-d6	5.721	84	394257	4.901	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	98.000%	
36) 1,2-Dichloropropane-d6	6.685	67	125575	4.987	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.800%	
41) Toluene-d8	7.894	98	331835	4.615	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	92.400%	
43) trans-1,3-Dichloroprop...	8.174	79	46733	4.786	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	95.800%	
46) 2-Hexanone-d5	8.627	63	179590	48.766	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.540%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	89916	4.936	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.800%	
66) 1,2-Dichlorobenzene-d4	12.190	152	111609	5.217	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.400%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121123\
Data File : VU056737.D
Acq On : 11 Dec 2023 17:17
Operator : MD/SY
Sample : 05728-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 12 01:11:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 12 01:07:35 2023
Response via : Initial Calibration

