

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121123\
 Data File : VU056763.D
 Acq On : 12 Dec 2023 04:11
 Operator : MD/SY
 Sample : 05675-13
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0AB6

Quant Time: Dec 12 05:05:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 12 02:01:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	215499	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	231298	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.808	152	91022	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	102357	5.541	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 110.800%		
7) Chloroethane-d5	1.908	69	83923	6.071	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 121.400%		
11) 1,1-Dichloroethene-d2	2.564	65	44632	5.647	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 113.000%		
20) 2-Butanone-d5	4.615	46	221322	60.671	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 121.340%		
24) Chloroform-d	5.055	84	208903	5.972	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 119.400%		
26) 1,2-Dichloroethane-d4	5.695	65	101776	6.149	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 123.000%		
32) Benzene-d6	5.721	84	403734	5.628	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 112.600%		
36) 1,2-Dichloropropane-d6	6.682	67	126312	5.625	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 112.600%		
41) Toluene-d8	7.895	98	332845	5.191	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 103.800%		
43) trans-1,3-Dichloroprop...	8.178	79	43398	4.984	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 99.600%		
46) 2-Hexanone-d5	8.628	63	170462	51.901	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 103.800%		
56) 1,1,2,2-Tetrachloroeth...	10.750	84	89121	5.485	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 109.800%		
66) 1,2-Dichlorobenzene-d4	12.187	152	109550	6.123	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 122.400%#		
Target Compounds						Qvalue
9) Trichlorofluoromethane	2.136	101	248900	8.010	ug/L	99
25) Chloroform	5.078	83	21937	0.614	ug/L	95
38) Bromodichloromethane	7.107	83	2876	0.108	ug/L #	77
47) Tetrachloroethene	8.550	164	29751	1.906	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121123\
Data File : VU056763.D
Acq On : 12 Dec 2023 04:11
Operator : MD/SY
Sample : 05675-13
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0AB6

Quant Time: Dec 12 05:05:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 12 02:01:50 2023
Response via : Initial Calibration

