

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121123\
 Data File : VU056753.D
 Acq On : 12 Dec 2023 00:09
 Operator : MD/SY
 Sample : VU1211WBL03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK096

Quant Time: Dec 12 02:02:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 12 02:01:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	253705	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	253613	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	107146	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	113287	5.209	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.200%
7) Chloroethane-d5	1.911	69	79343	4.876	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.600%
11) 1,1-Dichloroethene-d2	2.563	65	50194	5.395	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	107.800%
20) 2-Butanone-d5	4.621	46	232585	54.157	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.320%
24) Chloroform-d	5.055	84	217301	5.277	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%
26) 1,2-Dichloroethane-d4	5.695	65	105319	5.405	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.000%
32) Benzene-d6	5.721	84	420492	5.345	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.000%
36) 1,2-Dichloropropane-d6	6.685	67	132836	5.395	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.000%
41) Toluene-d8	7.894	98	355985	5.063	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
43) trans-1,3-Dichloroprop...	8.177	79	47817	5.008	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.200%
46) 2-Hexanone-d5	8.631	63	178853	49.665	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.320%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	93713	5.261	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.200%
66) 1,2-Dichlorobenzene-d4	12.190	152	117131	5.561	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121123\
Data File : VU056753.D
Acq On : 12 Dec 2023 00:09
Operator : MD/SY
Sample : VU1211WBL03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK096

Quant Time: Dec 12 02:02:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 12 02:01:50 2023
Response via : Initial Calibration

