

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U121219\
 Data File : VU036187.D
 Acq On : 13 Dec 2019 07:18
 Operator : JC/MD
 Sample : K6240-17
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFQX2

Quant Time: Dec 13 08:24:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121219WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 13 04:38:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.29	114	252573	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.45	117	271598	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.83	152	107327	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.61	65	101491	3.96	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.20%
7) Chloroethane-d5	1.93	69	101614	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.40%
11) 1,1-Dichloroethene-d2	2.59	63	106137	3.14	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	62.80%
20) 2-Butanone-d5	4.69	46	221510	53.87	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.74%
24) Chloroform-d	5.11	84	177073	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.20%
26) 1,2-Dichloroethane-d4	5.75	65	93686	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
32) Benzene-d6	5.77	84	331004	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.20%
36) 1,2-Dichloropropane-d6	6.73	67	107469	4.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.20%
41) Toluene-d8	7.93	98	260177	3.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.60%
43) trans-1,3-Dichloropropene-	8.21	79	33875	3.59	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	71.80%
46) 2-Hexanone-d5	8.67	63	160707	46.29	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.58%
57) 1,1,2,2-Tetrachloroethane-	10.78	84	99233	4.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	98.40%
64) 1,2-Dichlorobenzene-d4	12.21	152	106278	5.21	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.20%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl chloride	1.62	62	13881	0.411	ug/L #	81
8) Chloroethane	1.95	64	6266	0.302	ug/L	89
13) Acetone	2.67	43	42665	11.926	ug/L	100
15) Methyl Acetate	3.00	43	3747	0.491	ug/L	96
16) Methylene chloride	3.08	84	8526	0.320	ug/L	87
19) 1,1-Dichloroethane	3.92	63	37074	1.021	ug/L	100
21) 2-Butanone	4.78	43	5922	1.242	ug/L #	67
22) cis-1,2-Dichloroethene	4.72	96	23923	1.177	ug/L	89
29) 1,1,1-Trichloroethane	5.36	97	13853	0.416	ug/L	98
33) Benzene	5.82	78	39033	0.467	ug/L	100
42) Toluene	8.00	91	13562	0.153	ug/L	99
51) Chlorobenzene	9.47	112	61556	1.068	ug/L	99
53) m,p-Xylene	9.72	106	6082	0.180	ug/L	78
54) o-Xylene	10.13	106	4084	0.126	ug/L	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121219\
Data File : VU036187.D
Acq On : 13 Dec 2019 07:18
Operator : JC/MD
Sample : K6240-17
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFQX2

Quant Time: Dec 13 08:24:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121219WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 13 04:38:02 2019
Response via : Initial Calibration

