

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121220\
 Data File : VU041593.D
 Acq On : 13 Dec 2020 01:59
 Operator : SY/MD
 Sample : L5060-10
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B15

Quant Time: Dec 13 12:23:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Dec 13 11:26:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	133325	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	131053	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	70870	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	33203	3.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.80%
7) Chloroethane-d5	1.92	69	31839	3.95	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.00%
11) 1,1-Dichloroethene-d2	2.58	63	54681	2.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.00%#
20) 2-Butanone-d5	4.65	46	149187	49.83	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.66%
24) Chloroform-d	5.08	84	81947	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	5.72	65	48026	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.80%
32) Benzene-d6	5.74	84	162389	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
36) 1,2-Dichloropropane-d6	6.70	67	53638	4.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.20%
41) Toluene-d8	7.91	98	141621	4.18	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.60%
43) trans-1,3-Dichloropropene-	8.19	79	21269	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
46) 2-Hexanone-d5	8.64	63	122346	42.30	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.60%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	52254	4.89	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.80%
64) 1,2-Dichlorobenzene-d4	12.19	152	60951	4.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.80%

Target Compounds					Ovalue
13) Acetone	2.66	43	17954	9.427 ug/L	95
25) Chloroform	5.11	83	6380	0.362 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121220\
Data File : VU041593.D
Acq On : 13 Dec 2020 01:59
Operator : SY/MD
Sample : L5060-10
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B15

Quant Time: Dec 13 12:23:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun Dec 13 11:26:38 2020
Response via : Initial Calibration

