

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U121220\
 Data File : VU041598.D
 Acq On : 13 Dec 2020 03:55
 Operator : SY/MD
 Sample : L5060-15
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B34

Quant Time: Dec 13 12:40:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sun Dec 13 11:26:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	136783	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	127380	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	61839	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	33088	3.09	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.80%
7) Chloroethane-d5	1.92	69	32430	3.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.40%
11) 1,1-Dichloroethene-d2	2.58	63	54672	2.78	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.60%#
20) 2-Butanone-d5	4.66	46	147772	48.11	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.22%
24) Chloroform-d	5.08	84	80362	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.72	65	46831	4.51	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.20%
32) Benzene-d6	5.74	84	161820	4.44	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.80%
36) 1,2-Dichloropropane-d6	6.70	67	52014	4.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.00%
41) Toluene-d8	7.90	98	141529	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
43) trans-1,3-Dichloropropene-	8.18	79	20685	4.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.40%
46) 2-Hexanone-d5	8.64	63	117219	41.70	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.40%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	50135	4.83	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	54985	5.00	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.00%

Target Compounds					Ovalue
25) Chloroform	5.10	83	7754	0.429 ug/L	96
34) Trichloroethene	6.56	95	5161	0.504 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121220\
Data File : VU041598.D
Acq On : 13 Dec 2020 03:55
Operator : SY/MD
Sample : L5060-15
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B34

Quant Time: Dec 13 12:40:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sun Dec 13 11:26:38 2020
Response via : Initial Calibration

