

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121222\
 Data File : VU052350.D
 Acq On : 12 Dec 2022 14:13
 Operator : JC/MD
 Sample : N5922-06
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK001

Quant Time: Dec 13 01:03:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Dec 13 01:01:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	250105	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	254976	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	104386	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	106794	4.390	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.800%
7) Chloroethane-d5	1.916	69	91065	4.819	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.400%
11) 1,1-Dichloroethene-d2	2.568	65	42510	4.875	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	97.400%
20) 2-Butanone-d5	4.639	46	275330	62.437	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	124.880%
24) Chloroform-d	5.060	84	187594	5.248	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.000%
26) 1,2-Dichloroethane-d4	5.700	65	97399	5.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.800%
32) Benzene-d6	5.726	84	392563	5.186	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.800%
36) 1,2-Dichloropropane-d6	6.687	67	128630	5.491	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.800%
41) Toluene-d8	7.896	98	316523	4.622	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
43) trans-1,3-Dichloroprop...	8.176	79	41506	4.862	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.200%
46) 2-Hexanone-d5	8.632	63	195742	54.105	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.200%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	109683	5.400	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.000%
66) 1,2-Dichlorobenzene-d4	12.188	152	119488	5.979	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	119.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121222\
Data File : VU052350.D
Acq On : 12 Dec 2022 14:13
Operator : JC/MD
Sample : N5922-06
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK001

Quant Time: Dec 13 01:03:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Dec 13 01:01:45 2022
Response via : Initial Calibration

