

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121223\
 Data File : VU056791.D
 Acq On : 12 Dec 2023 17:20
 Operator : MD/SY
 Sample : 05664-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EOY69

Quant Time: Dec 13 00:34:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 13 00:30:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	239035	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	242349	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	98246	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	92640	4.521	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	90.400%	
7) Chloroethane-d5	1.907	69	78872	5.144	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	102.800%	
11) 1,1-Dichloroethene-d2	2.563	65	40769	4.651	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	93.000%	
20) 2-Butanone-d5	4.618	46	203923	50.397	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.800%	
24) Chloroform-d	5.055	84	195212	5.031	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	100.600%	
26) 1,2-Dichloroethane-d4	5.695	65	94945	5.171	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	103.400%	
32) Benzene-d6	5.721	84	363880	4.841	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6	6.685	67	114115	4.850	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	97.000%	
41) Toluene-d8	7.894	98	302438	4.502	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.000%	
43) trans-1,3-Dichloroprop...	8.177	79	43191	4.734	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	94.600%	
46) 2-Hexanone-d5	8.627	63	161300	46.872	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	93.740%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	82924	4.871	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	97.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	107528	5.568	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.400%	
Target Compounds						Qvalue
22) cis-1,2-Dichloroethene	4.663	96	8623	0.403	ug/L	89
42) Toluene	7.968	91	24945	0.283	ug/L	96
47) Tetrachloroethene	8.547	164	2140	0.131	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121223\
Data File : VU056791.D
Acq On : 12 Dec 2023 17:20
Operator : MD/SY
Sample : 05664-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EOY69

Quant Time: Dec 13 00:34:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 13 00:30:00 2023
Response via : Initial Calibration

