

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121223\
 Data File : VU056789.D
 Acq On : 12 Dec 2023 16:32
 Operator : MD/SY
 Sample : 05664-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 E0Y41

Quant Time: Dec 13 00:33:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 13 00:30:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	235022	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	230856	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	97966	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	93967	4.664	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.200%
7) Chloroethane-d5	1.907	69	72694	4.822	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.400%
11) 1,1-Dichloroethene-d2	2.563	65	42458	4.926	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	98.600%
20) 2-Butanone-d5	4.618	46	219477	55.167	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	110.340%
24) Chloroform-d	5.055	84	202013	5.296	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.000%
26) 1,2-Dichloroethane-d4	5.695	65	99072	5.488	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.800%
32) Benzene-d6	5.721	84	385848	5.389	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.800%
36) 1,2-Dichloropropane-d6	6.685	67	122911	5.484	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.600%
41) Toluene-d8	7.894	98	317034	4.954	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.000%
43) trans-1,3-Dichloroprop...	8.174	79	44224	5.088	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.800%
46) 2-Hexanone-d5	8.631	63	171046	52.179	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.360%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	87171	5.376	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	107.600%
66) 1,2-Dichlorobenzene-d4	12.187	152	107427	5.578	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.600%
Target Compounds						
					Qvalue	
65) 1,4-Dichlorobenzene	11.833	146	11452	0.318	ug/L	93
67) 1,2-Dichlorobenzene	12.206	146	9987	0.299	ug/L	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121223\
Data File : VU056789.D
Acq On : 12 Dec 2023 16:32
Operator : MD/SY
Sample : 05664-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
E0Y41

Quant Time: Dec 13 00:33:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 13 00:30:00 2023
Response via : Initial Calibration









#65
1,4-Dichlorobenzene
Concen: 0.318 ug/L
RT: 11.833 min Scan# 31
Delta R.T. 0.003 min
Lab File: VU056789.D
Acq: 12 Dec 2023 16:32

Instrument :
MSVOA_U
ClientSampleId :
E0Y41

Tgt Ion:146 Resp: 11452
Ion Ratio Lower Upper
146 100
111 50.1 31.2 58.0
148 60.9 45.6 84.8



#67
1,2-Dichlorobenzene
Concen: 0.299 ug/L
RT: 12.206 min Scan# 3395
Delta R.T. -0.003 min
Lab File: VU056789.D
Acq: 12 Dec 2023 16:32

Tgt Ion:146 Resp: 9987
Ion Ratio Lower Upper
146 100
111 44.2 36.9 55.3
148 65.9 53.0 79.4

