

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121318\
 Data File : VU028639.D
 Acq On : 12 Dec 2018 09:56
 Operator : MD/SY
 Sample : J6343-01
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BP-TB06-20181209

Quant Time: Dec 13 03:18:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	116183	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	208738	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	187746	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	75428	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	88381	49.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.70%	
35) Dibromofluoromethane	4.88	113	68483	52.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.26%	
50) Toluene-d8	7.56	98	267830	51.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.08%	
62) 4-Bromofluorobenzene	10.30	95	89152	45.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121318\
Data File : VU028639.D
Acq On : 12 Dec 2018 09:56
Operator : MD/SY
Sample : J6343-01
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BP-TB06-20181209

Quant Time: Dec 13 03:18:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

