

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121318\
 Data File : VU028656.D
 Acq On : 12 Dec 2018 16:47
 Operator : MD/SY
 Sample : J6343-08
 Misc : 5.0mL/MSVOA U/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BPS1-TT-MW311I-20181209

Quant Time: Dec 13 03:08:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	93121	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.88	114	167360	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	148647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	59715	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	73765	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
35) Dibromofluoromethane	4.88	113	56154	53.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.64%	
50) Toluene-d8	7.56	98	212810	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	10.30	95	71440	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
Target Compounds						
19) Methyl tert-butyl Ether	3.00	73	10965	2.790	ug/l	96
27) cis-1,2-Dichloroethene	4.22	96	9368	7.090	ug/l	87
30) Chloroform	4.67	83	2179	1.001	ug/l	95
44) Trichloroethene	6.19	130	1992	1.661	ug/l	84
52) Toluene	7.64	92	1577	0.497	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121318\
Data File : VU028656.D
Acq On : 12 Dec 2018 16:47
Operator : MD/SY
Sample : J6343-08
Misc : 5.0mL/MSVOA U/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BPS1-TT-MW311I-20181209

Quant Time: Dec 13 03:08:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

