

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121322\
 Data File : VU052387.D
 Acq On : 13 Dec 2022 19:05
 Operator : JC/MD
 Sample : N6030-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 COL33

Quant Time: Dec 14 03:22:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 14 03:17:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	235318	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	240008	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	105220	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	80540	3.519	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	70.400%	
7) Chloroethane-d5	1.916	69	76197	4.286	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	85.800%	
11) 1,1-Dichloroethene-d2	2.568	65	34011	4.145	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	83.000%	
20) 2-Butanone-d5	4.636	46	268674	64.756	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	129.520%	
24) Chloroform-d	5.060	84	166936	4.964	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.200%	
26) 1,2-Dichloroethane-d4	5.700	65	90207	5.310	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	106.200%	
32) Benzene-d6	5.726	84	339118	4.759	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.200%	
36) 1,2-Dichloropropane-d6	6.687	67	116003	5.261	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.200%	
41) Toluene-d8	7.896	98	272795	4.231	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	84.600%	
43) trans-1,3-Dichloroprop...	8.176	79	36059	4.487	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	89.800%	
46) 2-Hexanone-d5	8.633	63	202447	59.448	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	118.900%	
56) 1,1,2,2-Tetrachloroeth...	10.751	84	105455	5.516	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	110.400%	
66) 1,2-Dichlorobenzene-d4	12.189	152	112464	5.583	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.600%	
Target Compounds						
3) Chloromethane	1.520	50	35788	1.344	ug/L	99
12) 1,1-Dichloroethene	2.578	96	5341	0.313	ug/L #	1
19) 1,1-Dichloroethane	3.867	63	5091	0.154	ug/L	97
22) cis-1,2-Dichloroethene	4.665	96	224820	11.707	ug/L	98
29) 1,1,1-Trichloroethane	5.321	97	3690	0.139	ug/L #	89
34) Trichloroethene	6.539	95	1814394	93.131	ug/L	99
42) Toluene	7.967	91	51733	0.644	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121322\
Data File : VU052387.D
Acq On : 13 Dec 2022 19:05
Operator : JC/MD
Sample : N6030-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0L33

Quant Time: Dec 14 03:22:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 14 03:17:56 2022
Response via : Initial Calibration

