

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121322\
 Data File : VU052382.D
 Acq On : 13 Dec 2022 17:08
 Operator : JC/MD
 Sample : N6034-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGPL1

Quant Time: Dec 14 03:21:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 14 03:17:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	220912	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	228621	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	98635	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	78075	3.633	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.600%
7) Chloroethane-d5	1.916	69	72852	4.365	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.200%
11) 1,1-Dichloroethene-d2	2.568	65	33043	4.290	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	85.800%
20) 2-Butanone-d5	4.633	46	266751	68.485	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	136.980%#
24) Chloroform-d	5.060	84	162118	5.135	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.600%
26) 1,2-Dichloroethane-d4	5.700	65	86638	5.433	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.600%
32) Benzene-d6	5.726	84	328413	4.839	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
36) 1,2-Dichloropropane-d6	6.687	67	110600	5.265	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.400%
41) Toluene-d8	7.896	98	259168	4.220	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.400%
43) trans-1,3-Dichloroprop...	8.176	79	34685	4.531	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.600%
46) 2-Hexanone-d5	8.632	63	204282	62.975	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	125.940%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	103664	5.692	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	113.800%
66) 1,2-Dichlorobenzene-d4	12.188	152	102346	5.420	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.400%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.581	96	3927	0.245	ug/L	# 1
17) Methyl tert-butyl Ether	3.362	73	4161	0.116	ug/L	97
22) cis-1,2-Dichloroethene	4.668	96	119811	6.646	ug/L	99
25) Chloroform	5.089	83	14698	0.468	ug/L	96
34) Trichloroethene	6.539	95	119285	6.428	ug/L	99
47) Tetrachloroethene	8.549	164	362287	25.472	ug/L	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121322\
Data File : VU052382.D
Acq On : 13 Dec 2022 17:08
Operator : JC/MD
Sample : N6034-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGPL1

Quant Time: Dec 14 03:21:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 14 03:17:56 2022
Response via : Initial Calibration









#12
 1,1-Dichloroethene
 Concen: 0.245 ug/L
 RT: 2.581 min Scan# 386
 Delta R.T. 0.000 min
 Lab File: VU052382.D
 Acq: 13 Dec 2022 17:08




Instrument :
 MSVOA_U
 ClientSampleId :
 BGPL1


Tgt Ion: 96 Resp: 3927
 Ion Ratio Lower Upper
 96 100
 61 286.0 114.8 213.2#
 63 1535.6 93.7 174.1#



#17
 Methyl tert-butyl Ether
 Concen: 0.116 ug/L
 RT: 3.362 min Scan# 629
 Delta R.T. 0.000 min
 Lab File: VU052382.D
 Acq: 13 Dec 2022 17:08

Tgt Ion: 73 Resp: 4161
 Ion Ratio Lower Upper
 73 100
 43 16.7 13.2 19.8
 57 20.3 17.8 26.8









#22
 cis-1,2-Dichloroethene
 Concen: 6.646 ug/L
 RT: 4.668 min Scan# 1103
 Delta R.T. 0.003 min
 Lab File: VU052382.D
 Acq: 13 Dec 2022 17:08




Instrument :
 MSVOA_U
 ClientSampleId :
 BGPL1


Tgt Ion: 96 Resp: 119811
 Ion Ratio Lower Upper
 96 100
 61 122.7 85.1 158.1
 98 61.7 42.8 79.6



#25
 Chloroform
 Concen: 0.468 ug/L
 RT: 5.089 min Scan# 1166
 Delta R.T. 0.003 min
 Lab File: VU052382.D
 Acq: 13 Dec 2022 17:08

Tgt Ion: 83 Resp: 14698
 Ion Ratio Lower Upper
 83 100
 85 66.1 43.8 81.4









#34
Trichloroethene
Concen: 6.428 ug/L
RT: 6.539 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU052382.D
Acq: 13 Dec 2022 17:08

Instrument :
MSVOA_U
ClientSampleId :
BGPL1

Tgt Ion: 95 Resp: 119285
Ion Ratio Lower Upper
95 100
97 62.4 43.7 81.1
132 93.6 64.3 119.3
130 97.9 67.3 124.9



#47
Tetrachloroethene
Concen: 25.472 ug/L
RT: 8.549 min Scan# 2242
Delta R.T. -0.003 min
Lab File: VU052382.D
Acq: 13 Dec 2022 17:08

Tgt Ion: 164 Resp: 362287
Ion Ratio Lower Upper
164 100
129 95.7 67.4 125.2
131 92.6 67.1 124.7
166 125.8 89.7 166.5

