

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121418\
 Data File : VU028667.D
 Acq On : 13 Dec 2018 09:47
 Operator : MD/SY
 Sample : VU1214WBL01
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1214WBL01

Quant Time: Dec 14 05:43:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA U\METHOD\524U112618DW.M
 Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
 QLast Update : Tue Nov 27 01:03:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----	-----	-----	-----	-----	-----	-----
1) Fluorobenzene	5.74	96	15749	1.00	ug/l	0.00

System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.31	95	4667	0.81	ug/l	0.00
Spiked Amount 1.000			Recovery	=	81.00%	
68) 1,2-Dichlorobenzene-d4	11.86	152	4488	0.84	ug/l	0.00
Spiked Amount 1.000			Recovery	=	84.00%	


Target Compounds	Ovalue
-----	-----


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121418\
Data File : VU028667.D
Acq On : 13 Dec 2018 09:47
Operator : MD/SY
Sample : VU1214WBL01
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU1214WBL01

Quant Time: Dec 14 05:43:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA U\METHOD\524U112618DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Nov 27 01:03:50 2018
Response via : Initial Calibration





#1

Fluorobenzene

Concen: 1.000 ug/l

RT: 5.74 min Scan# 1420

Delta R.T. -0.00 min

Lab File: VU028667.D

Acq: 13 Dec 2018 09:47

Instrument :

MSVOA_U

ClientSampleId :

VU1214WBL01

Tot Ion: 96 Resp: 15749


Ion Ratio Lower Upper

96 100

70 22.4 17.8 26.8

95 9.3 7.6 11.4

97 5.7 0.0 0.0#






Abundance Ion 96.00 (95.70 to 96.70): VU0

Ion 70.00 (69.70 to 70.70): VU0

Ion 95.00 (94.70 to 95.70): VU0

Ion 97.00 (96.70 to 97.70): VU0



#57

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 10.31 min Scan# 2841

Delta R.T. 0.00 min

Lab File: VU028667.D

Acq: 13 Dec 2018 09:47


Tgt Ion: 95 Resp: 4667

Ion Ratio Lower Upper

95 100

174 74.0 59.3 88.9



176 74.4 54.3 81.5




Abundance Ion 95.00 (94.70 to 95.70): VU0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V






#68
 1,2-Dichlorobenzene-d4
 Concen: N.D. ug/l
 RT: 11.86 min Scan# 3324
 Delta R.T. 0.00 min
 Lab File: VU028667.D
 Acq: 13 Dec 2018 09:47

Instrument :
 MSVOA_U
 ClientSampleId :
 VU1214WBL01

Tot Ion:152 Resp: 4488
 Ion Ratio Lower Upper
 152 100
 115 63.3 0.0 146.8
 150 159.6 0.0 653.2



Abundance Ion 152.00 (151.70 to 152.70): V
 Ion 115.00 (114.70 to 115.70): V
 Ion 150.00 (149.70 to 150.70): V

