

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\U121420\
 Data File : VU041604.D
 Acq On : 14 Dec 2020 11:47
 Operator : SY/MD
 Sample : L5023-19
 Misc : 25.0mL/MSVOA U/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VHBLK01

Quant Time: Dec 14 12:49:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Dec 14 12:40:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.26	114	128585	5.00	ug/L	0.00
28) Chlorobenzene-d5	9.42	117	133317	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.81	152	64315	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.60	65	49933	4.97	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.40%
7) Chloroethane-d5	1.92	69	40669	5.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.60%
11) 1,1-Dichloroethene-d2	2.58	63	70987	3.83	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.60%
20) 2-Butanone-d5	4.66	46	156235	54.10	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.20%
24) Chloroform-d	5.08	84	87812	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
26) 1,2-Dichloroethane-d4	5.72	65	51305	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
32) Benzene-d6	5.74	84	179641	4.71	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.20%
36) 1,2-Dichloropropane-d6	6.70	67	56771	4.74	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.80%
41) Toluene-d8	7.90	98	156978	4.56	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.20%
43) trans-1,3-Dichloropropene-	8.19	79	25114	4.67	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.40%
46) 2-Hexanone-d5	8.64	63	137266	46.66	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.32%
57) 1,1,2,2-Tetrachloroethane-	10.75	84	51966	4.78	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.60%
64) 1,2-Dichlorobenzene-d4	12.19	152	58237	5.09	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.80%

Target Compounds					Ovalue
16) Methylene chloride	3.06	84	1508	0.145 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121420\
Data File : VU041604.D
Acq On : 14 Dec 2020 11:47
Operator : SY/MD
Sample : L5023-19
Misc : 25.0mL/MSVOA U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VHBLK01

Quant Time: Dec 14 12:49:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\SOMUTR121220WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Dec 14 12:40:03 2020
Response via : Initial Calibration

