

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121421\
 Data File : VU046348.D
 Acq On : 14 Dec 2021 12:20
 Operator : SY/MD
 Sample : VU1214WBL01
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: Dec 15 01:41:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 15 01:40:37 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	142479	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	150243	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	71929	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	52694	44.897	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.800%
7) Chloroethane-d5	1.919	69	41889	46.485	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.960%
11) 1,1-Dichloroethene-d2	2.575	63	70891	33.646	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.300%
21) 2-Butanone-d5	4.646	46	86292	92.488	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.490%
24) Chloroform-d	5.070	84	100330	51.190	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.380%
26) 1,2-Dichloroethane-d4	5.707	65	65857	50.068	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.140%
32) Benzene-d6	5.732	84	189911	44.101	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	6.694	67	60789	45.549	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.100%
41) Toluene-d8	7.903	98	161616	41.220	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.440%
43) trans-1,3-Dichloroprop...	8.182	79	28165	43.718	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.440%
47) 2-Hexanone-d5	8.636	63	61616	97.375	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.370%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	95470	47.170	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.340%
66) 1,2-Dichlorobenzene-d4	12.195	152	65124	47.013	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121421\
Data File : VU046348.D
Acq On : 14 Dec 2021 12:20
Operator : SY/MD
Sample : VU1214WBL01
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Dec 15 01:41:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 15 01:40:37 2021
Response via : Initial Calibration

