

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121422\
 Data File : VU052400.D
 Acq On : 14 Dec 2022 15:10
 Operator : JC/MD
 Sample : N6034-04
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGPL2

Quant Time: Dec 14 21:39:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 14 21:38:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	249019	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	258415	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	111142	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.600	65	75829	3.131	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.600%
7) Chloroethane-d5	1.915	69	72514	3.854	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	77.000%
11) 1,1-Dichloroethene-d2	2.568	65	31674	3.648	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.000%
20) 2-Butanone-d5	4.639	46	258052	58.774	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	117.540%
24) Chloroform-d	5.060	84	168220	4.727	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.700	65	87757	4.882	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.600%
32) Benzene-d6	5.726	84	328350	4.280	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.600%
36) 1,2-Dichloropropane-d6	6.687	67	112782	4.750	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.000%
41) Toluene-d8	7.896	98	257934	3.716	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.400%
43) trans-1,3-Dichloroprop...	8.176	79	34570	3.996	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.000%
46) 2-Hexanone-d5	8.632	63	183019	49.915	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.820%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	103175	5.012	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.200%
66) 1,2-Dichlorobenzene-d4	12.188	152	105639	4.965	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%
Target Compounds					Qvalue	
13) Acetone	2.655	43	57543	19.300	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121422\
Data File : VU052400.D
Acq On : 14 Dec 2022 15:10
Operator : JC/MD
Sample : N6034-04
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGPL2

Quant Time: Dec 14 21:39:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 14 21:38:14 2022
Response via : Initial Calibration

