

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121422\
 Data File : VU052397.D
 Acq On : 14 Dec 2022 14:00
 Operator : JC/MD
 Sample : N6031-03
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 C0B25

Quant Time: Dec 14 21:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 14 21:38:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	238937	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	245097	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	99594	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	82652	3.556	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	71.200%
7) Chloroethane-d5	1.916	69	77382	4.287	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.800%
11) 1,1-Dichloroethene-d2	2.568	65	33658	4.040	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.800%
20) 2-Butanone-d5	4.636	46	256799	60.957	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.920%
24) Chloroform-d	5.060	84	171293	5.016	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%
26) 1,2-Dichloroethane-d4	5.700	65	90246	5.232	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.600%
32) Benzene-d6	5.726	84	345814	4.752	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.000%
36) 1,2-Dichloropropane-d6	6.687	67	115391	5.124	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.400%
41) Toluene-d8	7.896	98	275118	4.179	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	8.176	79	35309	4.303	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.633	63	181042	52.059	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	104.120%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	103948	5.324	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.400%
66) 1,2-Dichlorobenzene-d4	12.189	152	104154	5.462	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	109.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121422\
Data File : VU052397.D
Acq On : 14 Dec 2022 14:00
Operator : JC/MD
Sample : N6031-03
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
C0B25

Quant Time: Dec 14 21:39:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 14 21:38:14 2022
Response via : Initial Calibration

