

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121422\
 Data File : VU052403.D
 Acq On : 14 Dec 2022 16:21
 Operator : JC/MD
 Sample : N6034-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGPJ9

Quant Time: Dec 14 21:40:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Dec 14 21:38:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	221871	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	230359	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	93824	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	67014	3.105	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	62.200%
7) Chloroethane-d5	1.916	69	63589	3.793	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.800%
11) 1,1-Dichloroethene-d2	2.568	65	28056	3.627	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.600%
20) 2-Butanone-d5	4.636	46	238175	60.884	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.760%
24) Chloroform-d	5.060	84	144211	4.548	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
26) 1,2-Dichloroethane-d4	5.700	65	78308	4.889	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	5.726	84	285390	4.173	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	6.687	67	98361	4.647	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.896	98	216674	3.502	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.000%
43) trans-1,3-Dichloroprop...	8.176	79	30255	3.923	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.400%
46) 2-Hexanone-d5	8.632	63	163201	49.931	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.860%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	93579	5.100	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	12.188	152	87560	4.874	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%
Target Compounds						Qvalue
17) Methyl tert-butyl Ether	3.359	73	8217	0.228	ug/L	98
22) cis-1,2-Dichloroethene	4.661	96	66378	3.666	ug/L	99
34) Trichloroethene	6.539	95	16353	0.875	ug/L	96
47) Tetrachloroethene	8.549	164	58987	4.116	ug/L	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121422\
Data File : VU052403.D
Acq On : 14 Dec 2022 16:21
Operator : JC/MD
Sample : N6034-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGPJ9

Quant Time: Dec 14 21:40:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Dec 14 21:38:14 2022
Response via : Initial Calibration









#17
Methyl tert-butyl Ether
Concen: 0.228 ug/L
RT: 3.359 min Scan# 61
Delta R.T. -0.003 min
Lab File: VU052403.D
Acq: 14 Dec 2022 16:21




Instrument :
MSVOA_U
ClientSampleId :
BGPJ9


Tgt Ion: 73 Resp: 8217
Ion Ratio Lower Upper
73 100
43 18.4 13.2 19.8
57 22.2 17.8 26.8



#22
cis-1,2-Dichloroethene
Concen: 3.666 ug/L
RT: 4.661 min Scan# 1033
Delta R.T. -0.003 min
Lab File: VU052403.D
Acq: 14 Dec 2022 16:21

Tgt Ion: 96 Resp: 66378
Ion Ratio Lower Upper
96 100
61 120.4 85.1 158.1
98 62.4 42.8 79.6





#34

Trichloroethene

Concen: 0.875 ug/L

RT: 6.539 min Scan# 1

Delta R.T. 0.000 min

Lab File: VU052403.D

Acq: 14 Dec 2022 16:21

Instrument :

MSVOA_U

ClientSampleId :

BGPJ9

Tgt Ion: 95 Resp: 16353





Ion Ratio Lower Upper

95 100

97 64.0 43.7 81.1

132 97.9 64.3 119.3

130 94.4 67.3 124.9



#47

Tetrachloroethene

Concen: 4.116 ug/L

RT: 8.549 min Scan# 2242

Delta R.T. 0.000 min

Lab File: VU052403.D

Acq: 14 Dec 2022 16:21

Tgt Ion: 164 Resp: 58987

Ion Ratio Lower Upper

164 100

129 96.7 67.4 125.2

131 90.4 67.1 124.7

166 121.3 89.7 166.5

