

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
 Data File : VU028689.D
 Acq On : 14 Dec 2018 17:39
 Operator : MD/SY
 Sample : J6393-03
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-9C_R1

Quant Time: Dec 15 01:08:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	76899	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	138562	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	124910	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	49161	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	65156	55.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.04%	
35) Dibromofluoromethane	4.88	113	47390	54.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.70%	
50) Toluene-d8	7.57	98	179025	51.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.80%	
62) 4-Bromofluorobenzene	10.30	95	57925	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.22	96	918	0.841	ug/l	96
44) Trichloroethene	6.19	130	918	0.925	ug/l	98
64) Tetrachloroethene	8.23	164	3012	3.615	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
Data File : VU028689.D
Acq On : 14 Dec 2018 17:39
Operator : MD/SY
Sample : J6393-03
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-9C_R1

Quant Time: Dec 15 01:08:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

