

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
 Data File : VU028690.D
 Acq On : 14 Dec 2018 18:03
 Operator : MD/SY
 Sample : J6393-04
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-9D_R1

Quant Time: Dec 15 01:09:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	77419	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	139241	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	126782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	48907	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	64910	54.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.88%	
35) Dibromofluoromethane	4.88	113	48292	55.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.22%	
50) Toluene-d8	7.57	98	179809	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
62) 4-Bromofluorobenzene	10.31	95	58378	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	819	0.746	ug/l	91
44) Trichloroethene	6.19	130	858	0.860	ug/l	92
64) Tetrachloroethene	8.23	164	2773	3.279	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
Data File : VU028690.D
Acq On : 14 Dec 2018 18:03
Operator : MD/SY
Sample : J6393-04
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-9D_R1

Quant Time: Dec 15 01:09:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

