

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
 Data File : VU028691.D
 Acq On : 14 Dec 2018 18:27
 Operator : MD/SY
 Sample : J6393-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-9E_R1

Manual Integrations
 APPROVED

MMDadoda
 12/17/2018 9:48:54 AM

Quant Time: Dec 15 01:12:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	78187	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	140904	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	126475	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	48984	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	65816	55.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.32%	
35) Dibromofluoromethane	4.88	113	47986	54.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.24%	
50) Toluene-d8	7.57	98	179568	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	10.30	95	58621	44.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.42%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.23	96	1005m	0.906	ug/l	
44) Trichloroethene	6.18	130	770	0.763	ug/l	82
64) Tetrachloroethene	8.23	164	2844	3.371	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
Data File : VU028691.D
Acq On : 14 Dec 2018 18:27
Operator : MD/SY
Sample : J6393-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-9E_R1

Manual Integrations
APPROVED

MMDadoda
12/17/2018 9:48:54 AM

Quant Time: Dec 15 01:12:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

