

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
 Data File : VU028692.D
 Acq On : 14 Dec 2018 18:51
 Operator : MD/SY
 Sample : J6393-06
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-10A_R1

Quant Time: Dec 15 01:15:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	75725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	138012	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	124113	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	49318	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	63129	54.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.26%	
35) Dibromofluoromethane	4.88	113	46804	53.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.78%	
50) Toluene-d8	7.57	98	176339	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
62) 4-Bromofluorobenzene	10.31	95	57931	45.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.22%	
Target Compounds						
21) trans-1,2-Dichloroethene	2.98	96	878	0.925	ug/l	96
27) cis-1,2-Dichloroethene	4.22	96	39417	36.686	ug/l	89
44) Trichloroethene	6.19	130	36455	36.860	ug/l	98
64) Tetrachloroethene	8.22	164	170260	205.662	ug/l	98
83) tert-Butylbenzene	11.10	119	4436	1.469	ug/l	79
85) sec-Butylbenzene	11.32	105	960	0.247	ug/l #	79
91) 1,2-Dichlorobenzene	11.88	146	2760	1.651	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
Data File : VU028692.D
Acq On : 14 Dec 2018 18:51
Operator : MD/SY
Sample : J6393-06
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-10A_R1

Quant Time: Dec 15 01:15:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

