

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
 Data File : VU028703.D
 Acq On : 14 Dec 2018 23:17
 Operator : MD/SY
 Sample : J6395-05
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 MW-1B_R1

Quant Time: Dec 15 02:43:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 07 00:39:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.98	168	75414	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.89	114	137001	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.09	117	124516	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	11.48	152	47676	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.31	65	65085	56.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.10%	
35) Dibromofluoromethane	4.88	113	47443	55.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.06%	
50) Toluene-d8	7.57	98	177884	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	10.31	95	57955	45.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.92%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1417	1.057	ug/l	92
27) cis-1,2-Dichloroethene	4.23	96	15233	14.236	ug/l	89
44) Trichloroethene	6.19	130	11855	12.075	ug/l	96
64) Tetrachloroethene	8.22	164	54185	65.240	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_U\DATA\VU121518\
Data File : VU028703.D
Acq On : 14 Dec 2018 23:17
Operator : MD/SY
Sample : J6395-05
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
MW-1B_R1

Quant Time: Dec 15 02:43:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_U\METHOD\82U120618W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 07 00:39:42 2018
Response via : Initial Calibration

