

Data Path : Z:\VOASRV\HPCHEM1\MSVOA U\DATA\VU121520\  
 Data File : VU041642.D  
 Acq On : 15 Dec 2020 17:32  
 Operator : SY/MD  
 Sample : L5029-04DL 250X  
 Misc : 25.0mL/MSVOA U/WATER  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_U  
 ClientSampleId :  
 GB1K4DL

Quant Time: Dec 16 05:49:33 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR121220WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Wed Dec 16 04:34:43 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 6.26  | 114  | 113962   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 9.42  | 117  | 115188   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.81 | 152  | 50418    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.60   | 65    | 28547    | 3.20     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 64.00%  |
| 7) Chloroethane-d5             | 1.92   | 69    | 31204    | 4.52     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 90.40%  |
| 11) 1,1-Dichloroethene-d2      | 2.58   | 63    | 49679    | 3.03     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 60.60%  |
| 20) 2-Butanone-d5              | 4.66   | 46    | 139988   | 54.70    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 109.40% |
| 24) Chloroform-d               | 5.08   | 84    | 75338    | 4.94     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 98.80%  |
| 26) 1,2-Dichloroethane-d4      | 5.72   | 65    | 45153    | 5.21     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 104.20% |
| 32) Benzene-d6                 | 5.74   | 84    | 147269   | 4.47     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 89.40%  |
| 36) 1,2-Dichloropropane-d6     | 6.70   | 67    | 47778    | 4.62     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 92.40%  |
| 41) Toluene-d8                 | 7.90   | 98    | 122955   | 4.13     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 82.60%  |
| 43) trans-1,3-Dichloropropene- | 8.18   | 79    | 19311    | 4.15     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 83.00%  |
| 46) 2-Hexanone-d5              | 8.64   | 63    | 110484   | 43.46    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 86.92%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.75  | 84    | 48393    | 5.15     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 103.00% |
| 64) 1,2-Dichlorobenzene-d4     | 12.19  | 152   | 49036    | 5.47     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 109.40% |

| Target Compounds      |      |     |       |            | Ovalue |
|-----------------------|------|-----|-------|------------|--------|
| 47) Tetrachloroethene | 8.56 | 164 | 45234 | 6.980 ug/L | 97     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA U\DATA\VU121520\  
Data File : VU041642.D  
Acq On : 15 Dec 2020 17:32  
Operator : SY/MD  
Sample : L5029-04DL 250X  
Misc : 25.0mL/MSVOA U/WATER  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_U  
ClientSampleId :  
GB1K4DL

Quant Time: Dec 16 05:49:33 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_U\METHOD\SOMUTR121220WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Wed Dec 16 04:34:43 2020  
Response via : Initial Calibration

