

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121521\
 Data File : VU046388.D
 Acq On : 15 Dec 2021 11:29
 Operator : SY/MD
 Sample : VU1215MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK064

Quant Time: Dec 16 03:08:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 16 02:25:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	149730	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	152211	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	78103	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	44198	35.834	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	71.660%
7) Chloroethane-d5	1.919	69	36462	38.503	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.000%
11) 1,1-Dichloroethene-d2	2.575	63	62909	28.412	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	56.820%#
21) 2-Butanone-d5	4.642	46	97467	99.406	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.410%
24) Chloroform-d	5.070	84	96890	47.041	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.080%
26) 1,2-Dichloroethane-d4	5.706	65	65805	47.605	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.220%
32) Benzene-d6	5.732	84	179977	41.254	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.500%
36) 1,2-Dichloropropane-d6	6.694	67	58890	43.556	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.120%
41) Toluene-d8	7.902	98	154833	38.980	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.960%#
43) trans-1,3-Dichloroprop...	8.182	79	27496	42.128	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.260%
47) 2-Hexanone-d5	8.636	63	70616	110.155	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	110.160%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	96515	47.070	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.140%
66) 1,2-Dichlorobenzene-d4	12.195	152	66236	44.036	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121521\
Data File : VU046388.D
Acq On : 15 Dec 2021 11:29
Operator : SY/MD
Sample : VU1215MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_U/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK064

Quant Time: Dec 16 03:08:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 16 02:25:11 2021
Response via : Initial Calibration

