

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121521\
 Data File : VU046392.D
 Acq On : 15 Dec 2021 13:14
 Operator : SY/MD
 Sample : M4983-22DL 5X
 Misc : 5.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 EW5S3DL

Quant Time: Dec 16 03:09:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Dec 16 02:25:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	145472	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.417	117	149122	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	73535	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	40122	33.482	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	66.960%	
7) Chloroethane-d5	1.919	69	32318	35.126	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	70.260%	
11) 1,1-Dichloroethene-d2	2.575	63	66568	30.944	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	61.880%	
21) 2-Butanone-d5	4.642	46	88322	92.716	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	92.720%	
24) Chloroform-d	5.067	84	89865	44.907	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	89.820%	
26) 1,2-Dichloroethane-d4	5.707	65	60718	45.211	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.420%	
32) Benzene-d6	5.732	84	167379	39.161	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	78.320%	
36) 1,2-Dichloropropane-d6	6.694	67	55823	42.143	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	84.280%	
41) Toluene-d8	7.899	98	139967	35.967	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	71.940%#	
43) trans-1,3-Dichloroprop...	8.179	79	26316	41.155	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	82.300%	
47) 2-Hexanone-d5	8.636	63	64320	102.412	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	102.410%	
56) 1,1,2,2-Tetrachloroeth...	10.758	84	91747	45.672	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	91.340%	
66) 1,2-Dichlorobenzene-d4	12.195	152	60065	42.414	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	84.820%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.588	96	20244	20.126	ug/L	86
17) trans-1,2-Dichloroethene	3.356	96	963	0.905	ug/L	89
19) 1,1-Dichloroethane	3.877	63	65203	33.689	ug/L	100
20) cis-1,2-Dichloroethene	4.671	96	19413	16.808	ug/L	100
30) 1,1,1-Trichloroethane	5.321	97	138802	75.828	ug/L	100
34) Trichloroethene	6.546	95	11498	9.670	ug/L	99
46) Tetrachloroethene	8.555	164	10064	11.518	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121521\
Data File : VU046392.D
Acq On : 15 Dec 2021 13:14
Operator : SY/MD
Sample : M4983-22DL 5X
Misc : 5.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
EW5S3DL

Quant Time: Dec 16 03:09:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Dec 16 02:25:11 2021
Response via : Initial Calibration

