

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121522\
 Data File : VU052414.D
 Acq On : 15 Dec 2022 12:25
 Operator : JC/MD
 Sample : N6050-01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 B0AA8

Quant Time: Dec 16 00:20:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 16 00:19:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	206609	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	218275	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	94950	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	25033	1.246	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	25.000%#
7) Chloroethane-d5	1.916	69	63813	4.088	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	81.800%
11) 1,1-Dichloroethene-d2	2.569	65	24334	3.378	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	67.600%
20) 2-Butanone-d5	4.642	46	258403	70.935	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	141.860%#
24) Chloroform-d	5.060	84	159309	5.395	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.000%
26) 1,2-Dichloroethane-d4	5.700	65	80653	5.408	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	108.200%
32) Benzene-d6	5.726	84	296768	4.580	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	6.687	67	103255	5.149	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.000%
41) Toluene-d8	7.896	98	169337	2.888	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	57.800%#
43) trans-1,3-Dichloroprop...	8.179	79	29116	3.984	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.633	63	172086	55.564	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	111.120%
56) 1,1,2,2-Tetrachloroeth...	10.752	84	100041	5.754	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	115.000%
66) 1,2-Dichlorobenzene-d4	12.189	152	95265	5.240	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.800%
Target Compounds						
					Qvalue	
25) Chloroform	5.086	83	130901	4.461	ug/L	95
38) Bromodichloromethane	7.102	83	67540	3.012	ug/L	96
49) Dibromochloromethane	8.806	129	30205	1.951	ug/L	96
59) Bromoform	10.292	173	1762	0.225	ug/L #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121522\
Data File : VU052414.D
Acq On : 15 Dec 2022 12:25
Operator : JC/MD
Sample : N6050-01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
B0AA8

Quant Time: Dec 16 00:20:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 16 00:19:11 2022
Response via : Initial Calibration

