

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121522\
 Data File : VU052420.D
 Acq On : 15 Dec 2022 16:51
 Operator : JC/MD
 Sample : N6034-17
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGPM7

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 12/19/2022
 Supervised By :Mahesh Dadoda 12/19/2022

Quant Time: Dec 16 00:21:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 16 00:19:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.247	114	199479	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.414	117	212733	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.809	152	89414	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	0.000	65	0d	0.000	ug/L	
Spiked Amount	5.000	Range	40 - 130	Recovery	=	0.000%#
7) Chloroethane-d5	1.916	69	59221	3.929	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.600%
11) 1,1-Dichloroethene-d2	2.568	65	3198	0.460	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	9.200%#
20) 2-Butanone-d5	4.639	46	238735	67.878	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	135.760%#
24) Chloroform-d	5.060	84	150515	5.279	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%
26) 1,2-Dichloroethane-d4	5.700	65	76353	5.302	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.000%
32) Benzene-d6	5.726	84	265727	4.207	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.687	67	92978	4.757	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.200%
41) Toluene-d8	7.896	98	1861	0.033	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	0.600%#
43) trans-1,3-Dichloroprop...	8.179	79	732m	0.103	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	2.000%#
46) 2-Hexanone-d5	8.632	63	147803	48.967	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.940%
56) 1,1,2,2-Tetrachloroeth...	10.751	84	92557	5.462	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	109.200%
66) 1,2-Dichlorobenzene-d4	12.188	152	86916	5.077	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%
Target Compounds						Qvalue
25) Chloroform	5.086	83	207276	7.317	ug/L	98
38) Bromodichloromethane	7.102	83	105660	4.835	ug/L	99
47) Tetrachloroethene	8.552	164	4929	0.372	ug/L	91
49) Dibromochloromethane	8.806	129	58236	3.860	ug/L	99
59) Bromoform	10.288	173	9430	1.280	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121522\
Data File : VU052420.D
Acq On : 15 Dec 2022 16:51
Operator : JC/MD
Sample : N6034-17
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGPM7

Quant Time: Dec 16 00:21:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 16 00:19:11 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 12/19/2022
Supervised By :Mahesh Dadoda 12/19/2022

