



Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121522\
 Data File : VU052409.D
 Acq On : 15 Dec 2022 08:14
 Operator : JC/MD
 Sample : BFB180
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BFB180

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR111822WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Fri Dec 16 00:19:11 2022



AutoFind: Scans 2889, 2890, 2891; Background Corrected with Scan 2876

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	17259	PASS
75	95	30	80	47.2	48856	PASS
95	95	100	100	100.0	103485	PASS
96	95	5	9	7.0	7245	PASS
173	174	0.00	2	1.4	1220	PASS
174	95	50	120	83.5	86445	PASS
175	174	5	9	7.2	6214	PASS
176	174	95	101	96.5	83384	PASS
177	176	5	9	7.0	5829	PASS

m/z	Abundance
36.00	1374.0
36.90	121.0
37.90	491.0
39.80	252.0
44.00	336.0

Instrument :
MSVOA_U
ClientSampleId :
BFB180

m/z	Abundance
36.00	2222.0
37.00	3923.0
38.00	3949.0
39.00	1739.0
39.90	282.0
44.00	1057.0
45.00	789.0
46.00	266.0
46.90	1076.0
47.90	532.0
49.00	4173.0
50.00	17360.0
51.10	5441.0
52.00	421.0
53.00	174.0
54.90	259.0
56.00	1308.0
57.00	2644.0
60.00	924.0
61.00	5072.0
62.00	5214.0
63.00	3894.0
64.00	952.0
65.10	1169.0
66.00	127.0
67.00	310.0
68.00	9541.0
69.00	10343.0
70.00	611.0
72.00	524.0
73.10	4606.0
74.00	15848.0
75.10	48920.0
76.00	4616.0
77.00	606.0
78.00	487.0
79.00	4203.0
79.90	1311.0
80.90	4157.0
81.90	911.0
82.80	139.0
86.00	205.0
87.00	5198.0
88.00	4802.0
90.90	362.0
92.00	2940.0
93.00	4136.0
94.00	11759.0
95.10	103088.0
96.10	7121.0
97.00	273.0
103.90	318.0
105.00	140.0
105.90	441.0
115.90	342.0
117.00	788.0
118.00	365.0
118.90	860.0
127.90	396.0
129.00	208.0
129.90	355.0
130.80	152.0
134.90	111.0
137.00	181.0
140.90	820.0
141.90	132.0
143.00	1765.0
144.90	877.0
145.80	161.0
147.00	125.0
147.90	247.0
149.80	100.0
155.00	357.0
156.90	140.0
171.00	328.0
172.00	1945.0
173.00	1190.0
174.00	83944.0
175.00	6224.0
176.00	81104.0
177.00	5794.0
177.90	129.0

Instrument :
MSVOA_U
ClientSampleId :
BFB180

m/z	Abundance
36.00	2248.0
37.00	3972.0
38.00	4283.0
39.00	1854.0
39.90	318.0
44.00	980.0
45.00	839.0
46.00	267.0
47.00	1047.0
47.90	467.0
49.00	4261.0
50.00	17840.0
51.10	5719.0
52.00	391.0
53.00	174.0
54.00	106.0
54.90	240.0
56.00	1313.0
57.00	2660.0
60.00	875.0
61.00	5229.0
62.00	5488.0
63.10	4041.0
64.00	967.0
65.10	1256.0
66.00	107.0
67.00	289.0
68.00	10122.0
69.00	10756.0
70.00	664.0
72.00	465.0
73.10	5053.0
74.10	16624.0
75.10	50496.0
76.00	4659.0
77.00	646.0
78.00	454.0
79.00	4181.0
79.90	1371.0
81.00	4198.0
82.00	967.0
82.80	149.0
86.00	191.0
87.00	5619.0
88.00	5332.0
90.90	334.0
92.00	2633.0
93.00	4393.0
94.00	12169.0
95.10	106824.0
96.00	7435.0
97.00	248.0
103.90	412.0
105.00	149.0
105.90	451.0
115.90	360.0
117.00	818.0
118.00	379.0
118.90	888.0
127.90	442.0
129.00	174.0
129.90	338.0
130.80	149.0
134.80	103.0
137.00	180.0
140.90	844.0
141.90	131.0
142.90	1793.0
143.90	107.0
144.90	841.0
145.80	146.0
147.00	137.0
147.90	243.0
149.80	124.0
155.00	396.0
156.90	170.0
171.00	396.0
172.00	1933.0
173.00	1291.0
174.00	89384.0
175.00	6365.0
176.00	85944.0
177.00	6080.0
177.90	135.0

Instrument :
MSVOA_U
ClientSampleId :
BFB180

m/z	Abundance
36.00	2070.0
37.00	3755.0
38.00	4110.0
39.00	1922.0
39.90	292.0
44.00	864.0
45.00	764.0
46.10	234.0
47.00	1019.0
48.00	415.0
49.00	3954.0
50.00	16576.0
51.10	5387.0
52.00	282.0
52.90	178.0
55.00	215.0
56.00	1297.0
57.00	2343.0
59.90	753.0
61.00	4801.0
62.00	4996.0
63.00	3810.0
64.00	883.0
65.10	1294.0
66.00	107.0
67.00	227.0
68.00	9626.0
69.00	9994.0
70.00	690.0
72.00	446.0
73.10	4714.0
74.10	15700.0
75.10	47152.0
76.00	4246.0
77.00	658.0
78.00	365.0
79.00	3817.0
79.90	1235.0
81.00	3974.0
81.90	996.0
82.90	106.0
85.90	177.0
87.00	5271.0
88.00	4884.0
90.90	279.0
92.00	2152.0
93.00	4475.0
94.00	11483.0
95.10	100544.0
96.00	7178.0
97.00	206.0
103.90	441.0
104.90	125.0
105.90	423.0
116.00	316.0
117.00	741.0
118.00	332.0
118.90	818.0
127.90	438.0
129.00	165.0
130.00	306.0
130.90	138.0
134.80	101.0
137.00	123.0
140.90	791.0
141.90	121.0
142.90	1630.0
143.90	121.0
145.00	788.0
145.80	127.0
147.00	122.0
147.90	205.0
149.80	123.0
155.00	328.0
156.90	232.0
158.90	100.0
161.00	105.0
171.00	423.0
172.00	1708.0
173.00	1180.0
174.00	86008.0
175.00	6052.0
176.00	83104.0
177.00	5612.0
178.00	125.0

Instrument :
MSVOA_U
ClientSampleId :
BFB180