

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
 Data File : VU056869.D
 Acq On : 15 Dec 2023 11:01
 Operator : MD/SY
 Sample : VU1215WBL01
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 VBLK102

Quant Time: Dec 15 22:24:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	244282	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	240839	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	99292	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	114369	5.462	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.200%
7) Chloroethane-d5	1.911	69	90108	5.751	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.000%
11) 1,1-Dichloroethene-d2	2.563	65	48324	5.394	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	107.800%
20) 2-Butanone-d5	4.628	46	219933	53.186	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.380%
24) Chloroform-d	5.055	84	215885	5.445	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.800%
26) 1,2-Dichloroethane-d4	5.695	65	103444	5.513	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.200%
32) Benzene-d6	5.721	84	414578	5.550	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.000%
36) 1,2-Dichloropropane-d6	6.685	67	129272	5.529	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	110.600%
41) Toluene-d8	7.894	98	338564	5.071	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.400%
43) trans-1,3-Dichloroprop...	8.177	79	44831	4.944	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	98.800%
46) 2-Hexanone-d5	8.631	63	169152	49.462	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.920%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	91268	5.395	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	108.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	109611	5.616	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
Data File : VU056869.D
Acq On : 15 Dec 2023 11:01
Operator : MD/SY
Sample : VU1215WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VBLK102

Quant Time: Dec 15 22:24:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration

