

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
 Data File : VU056875.D
 Acq On : 15 Dec 2023 13:33
 Operator : MD/SY
 Sample : 05677-07DL 5X
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 H0AA8DL

Quant Time: Dec 15 22:25:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	235171	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	238709	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	100427	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	94355	4.681	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.600%
7) Chloroethane-d5	1.911	69	76907	5.098	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.000%
11) 1,1-Dichloroethene-d2	2.563	65	41737	4.839	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.800%
20) 2-Butanone-d5	4.624	46	205679	51.666	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.340%
24) Chloroform-d	5.055	84	193995	5.082	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.600%
26) 1,2-Dichloroethane-d4	5.695	65	92286	5.109	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%
32) Benzene-d6	5.721	84	368248	4.974	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.400%
36) 1,2-Dichloropropane-d6	6.682	67	116050	5.008	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.200%
41) Toluene-d8	7.894	98	297560	4.497	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.000%
43) trans-1,3-Dichloroprop...	8.177	79	41944	4.667	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.400%
46) 2-Hexanone-d5	8.631	63	158864	46.868	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.740%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	83032	4.952	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.000%
66) 1,2-Dichlorobenzene-d4	12.187	152	105218	5.330	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.600%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.576	96	15821	0.829	ug/L	# 1
22) cis-1,2-Dichloroethene	4.660	96	6416	0.305	ug/L	85
34) Trichloroethene	6.534	95	124577	5.413	ug/L	97
47) Tetrachloroethene	8.550	164	119912	7.446	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
Data File : VU056875.D
Acq On : 15 Dec 2023 13:33
Operator : MD/SY
Sample : 05677-07DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
H0AA8DL

Quant Time: Dec 15 22:25:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration

