

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
 Data File : VU056901.D
 Acq On : 16 Dec 2023 00:28
 Operator : MD/SY
 Sample : 05809-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C30

Quant Time: Dec 16 05:12:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	205897	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	214684	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	88327	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.596	65	81155	4.598	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.907	69	62770	4.753	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.000%
11) 1,1-Dichloroethene-d2	2.560	65	36140	4.786	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.800%
20) 2-Butanone-d5	4.621	46	195188	56.002	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	112.000%
24) Chloroform-d	5.055	84	175120	5.240	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.800%
26) 1,2-Dichloroethane-d4	5.695	65	87395	5.526	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.600%
32) Benzene-d6	5.721	84	324683	4.876	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
36) 1,2-Dichloropropane-d6	6.685	67	107349	5.151	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.000%
41) Toluene-d8	7.894	98	259121	4.354	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.000%
43) trans-1,3-Dichloroprop...	8.177	79	35553	4.399	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.000%
46) 2-Hexanone-d5	8.631	63	150390	49.333	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.660%
56) 1,1,2,2-Tetrachloroeth...	10.750	84	78786	5.225	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	104.400%
66) 1,2-Dichlorobenzene-d4	12.190	152	93577	5.389	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
Data File : VU056901.D
Acq On : 16 Dec 2023 00:28
Operator : MD/SY
Sample : 05809-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C30

Quant Time: Dec 16 05:12:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration

