

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
 Data File : VU056904.D
 Acq On : 16 Dec 2023 01:41
 Operator : MD/SY
 Sample : 05809-11
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C51

Quant Time: Dec 16 05:13:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	203403	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.412	117	213382	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	87996	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	76803	4.405	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	88.200%	
7) Chloroethane-d5	1.911	69	63734	4.885	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	97.600%	
11) 1,1-Dichloroethene-d2	2.563	65	34168	4.580	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	91.600%	
20) 2-Butanone-d5	4.627	46	192491	55.905	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	111.820%	
24) Chloroform-d	5.055	84	168682	5.109	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.200%	
26) 1,2-Dichloroethane-d4	5.695	65	81879	5.241	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.800%	
32) Benzene-d6	5.721	84	310355	4.689	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	93.800%	
36) 1,2-Dichloropropane-d6	6.685	67	102060	4.927	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.600%	
41) Toluene-d8	7.894	98	247546	4.185	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	83.600%	
43) trans-1,3-Dichloroprop...	8.177	79	33948	4.226	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.600%	
46) 2-Hexanone-d5	8.630	63	139257	45.960	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	91.920%	
56) 1,1,2,2-Tetrachloroeth...	10.749	84	76002	5.071	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	101.400%	
66) 1,2-Dichlorobenzene-d4	12.190	152	92048	5.321	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.400%	
Target Compounds						
13) Acetone	2.634	43	81080	38.641	ug/L	Qvalue 99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121523\
Data File : VU056904.D
Acq On : 16 Dec 2023 01:41
Operator : MD/SY
Sample : 05809-11
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C51

Quant Time: Dec 16 05:13:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration





#13

Acetone

Concen: 38.641 ug/L

RT: 2.634 min Scan# 41

Delta R.T. 0.006 min

Lab File: VU056904.D


Acq: 16 Dec 2023 01:41

Instrument :

MSVOA_U

ClientSampleId :

F7C51



Tgt Ion: 43 Resp: 81080

Ion Ratio Lower Upper

43 100

58 35.8 0.0 72.6

