

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121621\
 Data File : VU046446.D
 Acq On : 17 Dec 2021 01:33
 Operator : SY/MD
 Sample : M5101-15ME 20X
 Misc : 6.30g/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGL39ME

Quant Time: Dec 17 04:06:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 17 03:50:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.250	114	138192	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	141537	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.813	152	70668	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.601	65	48131	42.281	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	84.560%		
7) Chloroethane-d5	1.919	69	36500	41.761	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	83.520%		
11) 1,1-Dichloroethene-d2	2.572	63	65617	32.109	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	64.220%		
21) 2-Butanone-d5	4.636	46	74247	82.047	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	82.050%		
24) Chloroform-d	5.067	84	90303	47.503	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	95.000%		
26) 1,2-Dichloroethane-d4	5.704	65	61575	48.265	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.520%		
32) Benzene-d6	5.729	84	178269	43.944	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	87.880%		
36) 1,2-Dichloropropane-d6	6.694	67	56392	44.854	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	89.700%		
41) Toluene-d8	7.900	98	154265	41.765	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	83.540%		
43) trans-1,3-Dichloroprop...	8.183	79	24999	41.190	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	82.380%		
47) 2-Hexanone-d5	8.636	63	55229	92.650	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	92.650%		
56) 1,1,2,2-Tetrachloroeth...	10.758	84	83240	43.658	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	87.320%		
66) 1,2-Dichlorobenzene-d4	12.195	152	61078	44.879	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	89.760%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121621\
Data File : VU046446.D
Acq On : 17 Dec 2021 01:33
Operator : SY/MD
Sample : M5101-15ME 20X
Misc : 6.30g/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGL39ME

Quant Time: Dec 17 04:06:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 17 03:50:47 2021
Response via : Initial Calibration

