

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121621\
 Data File : VU046452.D
 Acq On : 17 Dec 2021 03:48
 Operator : SY/MD
 Sample : M5121-07 20X
 Misc : 5.76g/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGL59

Quant Time: Dec 17 04:08:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 17 03:50:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	142619	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	147218	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	74520	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	49246	41.918	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	83.840%
7) Chloroethane-d5	1.919	69	35542	39.403	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.800%
11) 1,1-Dichloroethene-d2	2.572	63	64528	30.596	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.200%
21) 2-Butanone-d5	4.636	46	73612	78.820	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	78.820%
24) Chloroform-d	5.067	84	90056	45.903	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.800%
26) 1,2-Dichloroethane-d4	5.703	65	61712	46.870	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.740%
32) Benzene-d6	5.729	84	179165	42.461	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.920%
36) 1,2-Dichloropropane-d6	6.694	67	56558	43.250	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.500%
41) Toluene-d8	7.899	98	155620	40.506	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.020%
43) trans-1,3-Dichloroprop...	8.182	79	25441	40.301	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.600%
47) 2-Hexanone-d5	8.636	63	55789	89.978	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.980%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	83652	42.181	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.360%
66) 1,2-Dichlorobenzene-d4	12.195	152	62309	43.417	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121621\
Data File : VU046452.D
Acq On : 17 Dec 2021 03:48
Operator : SY/MD
Sample : M5121-07 20X
Misc : 5.76g/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGL59

Quant Time: Dec 17 04:08:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 17 03:50:47 2021
Response via : Initial Calibration

