

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121621\
 Data File : VU046458.D
 Acq On : 17 Dec 2021 06:04
 Operator : SY/MD
 Sample : M5121-14 20X
 Misc : 4.91g/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGL64

Quant Time: Dec 17 06:51:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 17 03:50:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.253	114	129259	50.000	ug/L	0.00
28) Chlorobenzene-d5	9.420	117	127980	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.812	152	64457	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.604	65	43646	40.991	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.980%
7) Chloroethane-d5	1.919	69	33830	41.381	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.760%
11) 1,1-Dichloroethene-d2	2.572	63	55378	28.972	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.940%#
21) 2-Butanone-d5	4.636	46	73761	87.143	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.140%
24) Chloroform-d	5.067	84	85939	48.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.660%
26) 1,2-Dichloroethane-d4	5.703	65	59663	49.998	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.000%
32) Benzene-d6	5.729	84	163719	44.633	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.260%
36) 1,2-Dichloropropane-d6	6.694	67	52874	46.511	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.020%
41) Toluene-d8	7.899	98	136071	40.742	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.480%
43) trans-1,3-Dichloroprop...	8.182	79	24164	44.032	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.060%
47) 2-Hexanone-d5	8.636	63	56061	104.008	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	104.010%
56) 1,1,2,2-Tetrachloroeth...	10.758	84	80932	46.944	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.880%
66) 1,2-Dichlorobenzene-d4	12.195	152	54373	43.802	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121621\
Data File : VU046458.D
Acq On : 17 Dec 2021 06:04
Operator : SY/MD
Sample : M5121-14 20X
Misc : 4.91g/5.0mL/100ul/5.0mL/MSVOA_U/MEOH
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGL64

Quant Time: Dec 17 06:51:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMULM112921WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 17 03:50:47 2021
Response via : Initial Calibration

