

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121622\
 Data File : VU052431.D
 Acq On : 16 Dec 2022 13:00
 Operator : JC/MD
 Sample : N6034-08
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGPJ9

Quant Time: Dec 16 21:12:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	16258	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	14874	0.500	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.742	152	7184	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	9856	0.514	ug/L	0.00
Spiked Amount 0.500	Range 40 - 130		Recovery = 102.000%			
4) 1,2-Dichloroethane-d4	5.661	65	8618	0.475	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 94.000%			
7) 1,2-Dichloropropane-d6	6.639	67	9795	0.548	ug/L	0.00
Spiked Amount 0.500	Range 60 - 140		Recovery = 110.000%			
8) Toluene-d8	7.830	98	12855	0.477	ug/L	0.00
Spiked Amount 0.500	Range 70 - 130		Recovery = 96.000%			
10) 1,1,2,2-Tetrachloroeth...	10.679	84	10547	0.552	ug/L	0.00
Spiked Amount 0.500	Range 65 - 120		Recovery = 110.000%			
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.605	62	2636	0.127	ug/L #	36
6) Trichloroethene	6.489	95	15656	0.940	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121622\
Data File : VU052431.D
Acq On : 16 Dec 2022 13:00
Operator : JC/MD
Sample : N6034-08
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGPJ9

Quant Time: Dec 16 21:12:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

