

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121622\
 Data File : VU052444.D
 Acq On : 16 Dec 2022 18:14
 Operator : JC/MD
 Sample : N6034-18
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 BGLP8

Quant Time: Dec 16 21:14:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	15922	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.346	117	14537	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.742	152	7201	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	12461	0.664	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 132.000%#	
4) 1,2-Dichloroethane-d4	5.661	65	11111	0.625	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 126.000%	
7) 1,2-Dichloropropane-d6	6.639	67	13130	0.752	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 150.000%#	
8) Toluene-d8	7.830	98	16209	0.616	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 124.000%	
10) 1,1,2,2-Tetrachloroeth...	10.679	84	13831	0.740	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 148.000%#	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121622\
Data File : VU052444.D
Acq On : 16 Dec 2022 18:14
Operator : JC/MD
Sample : N6034-18
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BGLP8

Quant Time: Dec 16 21:14:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

