

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121622\
 Data File : VU052441.D
 Acq On : 16 Dec 2022 17:04
 Operator : JC/MD
 Sample : N6034-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 20 Sample Multiplier: 1

Quant Time: Dec 16 21:13:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Dec 02 00:27:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.206	114	17242	0.500	ug/L	0.00
5) Chlorobenzene-d5	9.361	117	15868	0.500	ug/L	0.01
11) 1,4-Dichlorobenzene-d4	11.741	152	7717	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.598	65	10565	0.520	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	104.000%
4) 1,2-Dichloroethane-d4	5.660	65	9740	0.506	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
7) 1,2-Dichloropropane-d6	6.639	67	11268	0.591	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	118.000%
8) Toluene-d8	7.830	98	14120	0.491	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.679	84	11944	0.586	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	118.000%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121622\
Data File : VU052441.D
Acq On : 16 Dec 2022 17:04
Operator : JC/MD
Sample : N6034-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 20 Sample Multiplier: 1

Quant Time: Dec 16 21:13:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUSIM120122.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Dec 02 00:27:38 2022
Response via : Initial Calibration

