

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
 Data File : VU056938.D
 Acq On : 16 Dec 2023 19:41
 Operator : MD/SY
 Sample : 05810-15
 Misc : 25.0mL/MSVOA_U/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_U
 ClientSampleId :
 F7C88

Quant Time: Dec 16 23:48:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Dec 15 01:53:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.245	114	193439	5.000	ug/L	0.00
28) Chlorobenzene-d5	9.415	117	204687	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.807	152	80363	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.595	65	66005	3.981	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	79.600%
7) Chloroethane-d5	1.907	69	54552	4.397	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	2.560	65	29480	4.155	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.200%
20) 2-Butanone-d5	4.621	46	186870	57.068	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	114.140%
24) Chloroform-d	5.055	84	157850	5.027	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	5.698	65	79638	5.360	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.200%
32) Benzene-d6	5.721	84	287380	4.527	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
36) 1,2-Dichloropropane-d6	6.685	67	94183	4.740	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.800%
41) Toluene-d8	7.894	98	225198	3.969	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.400%
43) trans-1,3-Dichloroprop...	8.177	79	31763	4.122	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.400%
46) 2-Hexanone-d5	8.631	63	139377	47.954	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.900%
56) 1,1,2,2-Tetrachloroeth...	10.753	84	74027	5.149	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.000%
66) 1,2-Dichlorobenzene-d4	12.190	152	78314	4.957	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.200%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.570	96	4246	0.270	ug/L	# 1
17) Methyl tert-butyl Ether	3.354	73	638297	18.330	ug/L	99
19) 1,1-Dichloroethane	3.859	63	218739	7.142	ug/L	98
22) cis-1,2-Dichloroethene	4.663	96	10160	0.587	ug/L	96
25) Chloroform	5.081	83	24266	0.756	ug/L	96
34) Trichloroethene	6.537	95	18299	0.927	ug/L	92
47) Tetrachloroethene	8.550	164	4961	0.359	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU121623\
Data File : VU056938.D
Acq On : 16 Dec 2023 19:41
Operator : MD/SY
Sample : 05810-15
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
F7C88

Quant Time: Dec 16 23:48:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\SFAMUTR121123WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Dec 15 01:53:53 2023
Response via : Initial Calibration

